

February 21th, 2025**Product code / Código de producto:** SWE 2.7

Product description / Descripción de producto: Chemdye® SWE high-absorption swabs have been specifically designed for use with cannulated instruments. Chemdye® SWE swabs should be used with the Chemdye® PRO1 ENDO system, which enables verification of the cleaning of cannulated instruments by detecting protein residues that may remain after the process. / Los hisopos Chemdye® SWE de alta absorción han sido diseñados específicamente para su uso en instrumentos canulados. Los hisopos Chemdye® SWE deben utilizarse junto con el sistema Chemdye® PRO1 ENDO, que permite verificar la limpieza de los instrumentos canulados mediante la detección de residuos de proteínas que puedan quedar después del proceso.

Batch number / Número de lote: C50009**Manufacture date / Fecha de fabricación:** 2025-02-01**Expiration date / Fecha de vencimiento*:** 2027-01-31

*When stored at temperatures between 15-30 °C / Cuando se almacenan a temperaturas entre 15-30 °C.

CONCLUSION: The product batch **C50009** meets the specifications of ISO 13485:2016/NS-EN ISO 13485:2016 standards. / El lote de producto **C50009** cumple con las especificaciones establecidas en las normativas ISO 13485:2016/NS-EN ISO 13485:2016.


TERRAGENE S.A.
Lic. Adrián Rovetto
DIRECTOR TÉCNICO